DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 **CONCERNING MEDICAL DEVICES**

MANUFACTURER: HETAIDA TECHNOLOGY CO., LTD

4F, BAISHIDA HIGH-TECH PARK, XIANGDONG INDUSTRIAL AREA, DALINGSHAN TOWN, DONGGUAN CITY,

GUANGDONG, CHINA.

MEDICAL DEVICE: Non-contact Infrared Body Thermometer MODEL: HTD8808C, HTD8818A, HTD8816C, HTD8813 \ HTD8819

CLASSIFICATION - ANNEX IX: CLASS IIA, RULE 10

CONFORMITY ASSESSMENT ROUTE: ANNEX II (excluding Section 4)

WE, HETAIDA TECHNOLOGY CO., LTD., AT OUR SOLE RESPONSIBLE HEREWITH DECLARE THAT

BODY MODE OF THE STATED MEDICAL DEVICES

MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF

14 JUNE 1993 CONCERNING MEDICAL DEVICES;

INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER.

AND are in conformity with the national standards transposing harmonised standards:

EN ISO 13485: :2016; EN ISO 14971: 2012; EN 1041:2008+A1: 2013; EN 62304:2006+A1:2015;

EN 60601-1-6:2010+A1:2015; EN 62366-1:2015;

EN 60601-1:2006+A1:2013; EN 60601-1-2:2015; EN 60601-1-11: 2015; EN ISO 80601-2-56:2017

NOTIFIED BODY: SGS FIMKO OY

Takomotie 8 00380 HELSINKI COUNTRY: Finland

Phone: +358 9 696 361 Fax: +358 9 692 5474

> E-MAIL: sgs.fimko@sgs.com Website: www.fi.sgs.com

IDENTIFICATION NUMBER

(EC) CERTIFICATE(S): FI17/07005

CE,598

EUROPEAN REPRESENTATIVE: WELLKANG LTD

EC REP

16 Castle St. Dover, CT16 1PW, UK TEL: +44(20)30869438 FAX: +44(20)76811874

START OF CE-MARKING: ISSUE DATE OF EC CERT

PLACE, DATE OF DECLARATION: CHINA GUANGDONG PROVINCE 2020.03.19

Tomchen

SIGNATURE:

NAME: TOM CHEN 陈振光

POSITION: (RESPONSIBLE SENIOR EXECUTIVE OF MANUFACTURER)

File No: ZHTF-CE-01-030 REV:A.2 Revision date: 2020.03.25